

Malaviya National Institute of Technology, Jaipur

J.L.N. MARG, JAIPUR-302017

Tel : (0141)2713312 (o)

Fax : (0141) 2529029

Bhurendra Singh

Deputy Registrar

CORRIGENDUM

Tender No. F5(2513)ST/MNIT/PHY/2019

Name of Lab :Electronics Lab

A pre-bid meeting was held on 16.01.2020 to address the queries of the tenders. After discussion department has finalised below mentioned modifications in the technical specifications of the equipment of above-mentioned tender. Below mentioned revisions may please be considered. For details visit: www.mnit.ac.in, <https://eprocure.gov.in/epublish/app> & <http://tenderwizard.com/MNITJ>

Item No.	Name of Item	Revised Specifications	Qty
12	Digital Oscilloscope	a. Sampling rate- 1GSa/Sec or higher b. The manufacturer must have NABL accredited service/calibration lab in India.	05 Nos
13	Function Generator	a. Frequency range 10 MHz must be read as 20 MHz or higher b. Number of channels- 2 c. Must have arbitrary waveform generation functionality d. The manufacturer must have NABL accredited service/calibration lab in India.	05 Nos

Note: Other specifications and conditions shall remain same as in the original tender

Deputy Registrar
(Store & Purchase)