

Bhurendra Singh

Deputy Registrar

CORRIGENDUM

Tender No. F5(1012)ST/MNIT/CEE/2019

A pre-bid meeting is held at the scheduled date and time i.e., 20.02.2020 (2:00 PM onwards) to address the queries of the tenders. After discussion department has finalised below mentioned modifications in the technical specifications of the equipment of above-mentioned tender. Below mentioned revisions may please be considered.

Sl. No.	Description	Tender Specification	Revised Specifications
1	Point number 3	Inconel 625	Inconel 625/ Inconel 600
2	Point number 7	Heating Elements: APM-Kanthal Heating Element Coil	Heating Elements: APM- Kanthal Heating Element Coil or appropriate technically suitable heating element for achieving required 1200 deg C
3	Point Number 13	Total assembly will be provided on a main common frame with hydraulic operating mechanism for charging and discharging arrangement.	Total assembly will be provided on a main common frame with hydraulic operating mechanism for charging and discharging arrangement or appropriate technically suitable heating element for achieving required 1200 deg C

Note: Other specifications and conditions shall remain same as in the original tender

Deputy Registrar
(Store & Purchase)