

Malaviya National Institute of Technology Jaipur

Jawahar Lal Nehru Marg, JAIPUR-302017 (Rajasthan)

Ministry of Education

(Government of India)



Global Tender Enquiry

For

Procurement of “Powder X-Ray Diffraction facility with residual stress measurement attachments”

MALAVIYA NATIONAL INSTITUTE OF TECHNOLOGY JAIPUR

No. F5(2736)ST/MNIT/Mett./2022

Phone : 0141-2713312,2713352

NOTICE INVITING GLOBAL TENDER

Registrar, MNIT, JLN Marg, Jaipur invites sealed tenders for the supply of **“Powder X-Ray Diffraction facility with residual stress measurement attachments”** for Met. & Mat. Engg. Deptt. of this Institute in **Two Bid System (Technical & Financial bids in separate envelope)** as per schedule given below:

Schedule	
Event	Date & Time
Download of Tender	13.01.2023
Bid Submission Last Date & Time	13.02.2023 by 2.00 PM
Technical Bid Opening Date & Time	13.02.2023 at 3.00 PM
Earnest Money	(Rs. 3,00,000.00) EMD in the form of Demand Draft in the name of The Registrar, MNIT payable at Jaipur or NEFT/RTGS in the bank account details as under:- Bank Name: ICICI Bank Ltd., Bank Branch: MNIT Jaipur Account Holder: Registrar MNIT Jaipur, J.L.N. Marg, Jaipur Bank Account No: 676805000011 Bank IFSC Code: ICIC0006768
Performance Bank Guarantee	3% of the contract value valid till warranty period plus 60 days.
Warranty	03 years
Delivery	90 Days
No. of Covers	02
Bid Validity days	120 days
Email Address (for Technical Clarifications)	storepurchase@mnit.ac.in

Quotation must be enclosed in a properly sealed envelope addressed to **Deputy Registrar (S&P), Malaviya National Institute of Technology, Jawahar Lal Nehru Marg, Jaipur -302017** (E-mail address storepurchase@mnit.ac.in) by designation and not by name. The quotations must be super scribed “Quotations for the supply of **“Powder X-Ray Diffraction facility with residual stress measurement attachments”** as called for in Tender Notice No. ----- dated _____ due on----- at ----- --AM/PM. The Quotation must reach on or before the due date and time mentioned in the ‘Schedule’. The documents must be dropped in the tender box available in Central Store during office hours (9.30am to 6.00pm) on all working days. Bids delivered to any other place or any individual shall not be considered as valid document. Quotations sent by e-mail will also not be considered valid. The complete Tender document can be viewed and downloaded only from the website (www.mnit.ac.in) and CPPP site <https://eprocure.gov.in/epublish/app> during the tender period.

2. The bids (complete in all respect) must be submitted in **two separate** Envelopes as explained below.

Envelope – 1 (Following documents to be provided)			
Sl. No.	Type	Content	Supporting Page Number of bid
1.	Technical Bid	Index / Table of Content	
2.		Proof of Earnest money Deposit as mentioned in the ‘Schedule’ . In case of exemption from submission of Earnest Money Deposit, proof of registration with NSIC/MSME.	
3.		Copy of GST Certificate and PAN.	
4.		Leaflets Catalogue (Brochure)	
5.		Compliance Certificate (Annexure-I)	
6.		Non Blacklisting of Supplier and Price reasonability declaration as per Annexure-II	
7.		Required past Experience as per Annexure- III (kindly refer clause No. 4.7.5)	
8.		Minimum average annual turnover of the bidder (kindly refer clause No. 4.7.6)	
9.		After Sale Service Certificate (Annexure IV) (kindly refer clause No. 4.30)	
10.		Manufacturer’s Authorization Letter as per Annexure-V	
11.		Certificate - Bidder Not from Country sharing Land border with India & Registration of Bidder with Competent Authority (Annexure-VII)	
12.		Declaration of Local Content (Annexure-VIII)	
Envelope - 2			
Sl. No.	Type	Content	
1.	Financial Bid	Price bid in specified format only. (Annexure- IX) <i>Note:</i> 1. Price is to be quoted in Indian Rupees only. 2.-Comparison of prices will be done ONLY on the bids submitted for the Main Equipment and anything asked as ‘Optional’ in the specifications is not to be included for overall comparison)	

Note: Price are to be quoted in Financial bid only placed in separate sealed envelope. If price are quoted anywhere in technical bid documents, the bid will summarily be rejected.

3. INVITATION FOR TENDER OFFER FOR PROCUREMENT OF “POWDER X-RAY DIFFRACTION WITH RESIDUAL STRESS MEASUREMENT FACILITY QUANTITY: 01”

Malaviya National Institute of Technology, Jaipur (MNIT) invites bids (Technical bids and Commercial bid) from eligible and experienced OEM (Original Equipment Manufacturer) or OEM Authorized Dealer for procurement of “Programmable Powder X-Ray Diffraction facility with residual stress measurement attachments /Qty., 01 (warranty period as stated at “Schedule”) on site comprehensive warranty from the date of receipt of the material as per terms & conditions specified in the tender document.

3.1. TECHNICAL SPECIFICATION:

Powder X-ray Diffraction with Residual Stress Measurement Facility/Quantity: 01

Name of the Item : Powder X-ray Diffraction with Residual Stress Measurement Facility/ Quantity: 1 No.			
Sl No.	Particulars	Comply (Yes/No)	Supporting Page No.
	Floor mounted fully automated X-Ray Diffraction (XRD) system along with all necessary hardware and software for analysis of powder and solid samples of metals/alloys/ceramic/polymer/composites. In future the system must have capability to be upgradable for SAXS, Microdiffraction, Texture and In-Situ Non-Ambient/high temperature Applications.		
1.	X-Ray Generator	a) Output power: 3 kW or more	
		b) Voltage: up to 60kV or higher with increment of 1 kV	
		c) Anode Current: up to 50mA or higher with increment of 1 mA	
		d) Control: Fully automated through software	
		e) Stability: ≤ 0.01 % per 10 % variation of main voltage or better at 50 Hz	
2.	Goniometer	a) Type: Higher resolution Theta/Theta	
		b) Scan: θ/θ coupled and θ/θ decoupled. Goniometer should be with independent θ and 2θ drives for accurate positioning and with design to have zero backlash.	
		c) Two theta range: from -90^0 to 162^0 or better range	
		d) Minimum step size: 0.0001^0 or better	
		e) Angular speed: 10-15 degree/sec or better	
		f) Angular resolution: 0.028 deg. or better	
		g) Goniometer radius: 220 mm or more. Suitable hardware and software to be included for goniometer radius change facility.	
3.	X-Ray Tube	a) Target: Copper (Cu)	
		b) Focus: Long fine focus (LFF) X-ray tube with 1 line and 1 point focus.	
		c) Insulation: Ceramic Insulation.	
		d) Maximum output power: ≥ 1.8 kW	
		e) Suitable Beta filter to be included.	
		f) X-ray tube must have auto component recognition facility to indicate anode target type and its exchange from one source type to another type.	
		g) Changeover from line to point focus and Vice-Versa should be available.	
		h) Alignment of X-Ray tube Height and Focus must be through computer control.	
4.	Sample stage and Sample holder	a) Standard spinning/rotating sample stage(s) suitable for reflection and transmission mode diffraction with facility to keep powder sample always in horizontal position.	
		b) For solid and thin film sample, a suitable sample stage to be offered with computer controlled motorized Z stage with up to 25 mm or more movement and capability to mount sample up to 2kg weight. The sample stage must be able to handle sample of minimum 100 mm diameter and 50 mm thickness	

5.	Optics	All optical components and its settings must be automatically recognized, changed and aligned by XRD control software for error free analysis with latest design.		
		a) Primary Beam Optics: Computer controlled motorized divergence slits, two numbers of suitable sollerslit/ s, motorized mask and beta filters to be offered. Multilayer mirror for Cu X-ray source to be offered for thin film and transmission. Changeover from Bragg Brentano geometry to parallel beam geometry must be computer controlled.		
		b) Secondary Beam: Computer controlled motorized Anti-scatter slits, suitable soller slit/s, long soller slit/parallel plate collimator to be offered for parallel beam geometry.		
		c) Filters: Suitable $K\beta$ -filters for Cu radiation		
		d) Monochromator to reduce the fluorescence from the sample.		
		e) The Line optics must include X-ray Mirror to carry out, thin film GIXRD.		
		f) For carrying out residual stress analysis, point focus of the x-ray beam must be useful.		
6.	X-ray Detector	The maintenance free solid state (semiconductor) detector should have following specifications;		
		a) High degree of linearity (i.e., 95% in the range of 0 to 10^8 CPS)		
		b) Very less background noise		
		c) 1D detector should be capable of working in 0D and 1D mode.		
		d) Number of active channels / pixels 190 or higher.		
		e) Spatial resolution should be 75 micron or less.		
		f) Detector should work in continuous, fast scan, slow scan and static modes.		
		g) It should also work in fluorescence reduction mode.		
		h) The detector should be able to work with most of the X-ray source like Cu, Mo, Co etc.		
		i) The detector should not use any liquid or gas in its operation.		
7.	Standards	a) Fluorescent disc/screen for positioning the X-ray beam		
		b) Silicon and/or other standard sample for checking the accuracy of the peak position within + 0.01 linearity range over the entire goniometer angular range.		
		c) Reference standard for Residual Stress Application		
8.	Software	a) Data collection - Licensed version of the software to control the complete XRD instrument, auto component recognition of all the components, system diagnosis and programming for data collection to be offered.		
		b) Data analysis - Licensed version of data evaluation software for background correction, profile fitting, search match, qualitative analysis, structure refinement, lattice parameter refinement, crystallite size determination, Rietveld based standardless quantification refinement package with Structure analysis, profile fitting, lattice parameter refinement, stripping of $K\alpha_2$ peak, and crystallite size determination software package should be offered. Qualitative and quantitative analysis software, search match, dedicated SAXS application software for analysis of bulk and nanoparticle has to be offered.		
		c) Latest version of ICDD PDF4+ AXIOM Database with minimum 3 years validity must be quoted.		
		d) Software for data collection and data analysis of thin film GIXRD analysis and residual stress analysis to be offered.		
		e) Data analysis software should be with minimum 3 user licenses for use by multiple scientists, faculties and researchers.		

9	XRD-analysis Computer	Suitable high end computer with all the software (as mentioned in point 8), 23inch monitor and a device to produce colored XRD patterns.		
10.	Safety:	a) Provision for detection of abnormal cooling water flow, pressure and temperature, abnormal generator overload and shutter malfunction detection.		
		b) XRD system should be compliant as per International Safety Standards and Regulation norms like Eurotom, pertaining mechanical, electrical and X-ray Radiation safety and other hazards.		
		c) The supplier must have necessary Type approval from AERB for supply of XRD and necessary approval document to be submitted along with offer.		
11	Cooling system:	a) A suitable external water chiller compatible with the main instrument whose Make and detailed specifications to be provided with the installation requirements. They should preferably be domestically procured.		
		b) External filter/traps should be provided for minimizing the dust settlements from the chiller to the target area.		
12	Power backup system for XRD	a) Minimum 15 kVA or more IGBT based online power backup with built-in isolation transformer.		
		b) Power back must be able to handle XRD system, XRD-analysis Computer, and external water chiller pump for more than 30 min uninterrupted XRD data collection during main power failure.		
13	Other mandatory requirements	a) List of recent installations (in last 3 years) of the offered XRD model in India should be provided.		
		b) After sales service will be a component for technical evaluation of the equipment. For this, list of service centers in north region with contact details of engineers must be submitted.		
		c) XRD system should be supported by manufacturer for the supply of spares for at least 8-10 years after Installation.		
		d) All pre-installation requirements such as room size, mechanical vibration and required power rating for the XRD system and water chiller are to be provided. The supplier must arrange site inspection for installation planning. The vendor should provide all required accessories and tackle for complete installation and commissioning of instrument.		
		e) All the operational manual, application manual and service manual along with schematic in English are to be provided. Test Reports for all the modes of operation to be provided.		
14	Application and Operation Training:	a) Supplier must provide complete operation, DO'S AND DON'TS, and application training for all aforesaid applications for minimum 5 days.		
		b) All costs towards training including lodging, boarding and travel have to be borne by the XRD supplier.		
15	Warranty:	a) XRD system including all accessories such as X-Ray Generator, X-ray tube, electronics, motors and detector must be provided with complete & comprehensive warranty of 3 years .		
		b) During warranty period, if any part is required for replacement, then same to be arranged on delivered duty paid (DDP) basis without any cost to buyer.		
		c) X-ray tube should be supplied before expiry of 3 rd year. This is to make sure that the XRD system will work without interruption for 5 years.		
Optional Item				
1	Accessories for residual stress and texture measurement	a) 3 Axis Elurian cradle with Chi and Phi movement with automatic computer control to be included.		
		b) Advance polycapillary lens to have point focus beam on sample for stress and texture measurement.		
		c) Sample holder adaptor for cradle must be provided.		
		d) Necessary software for texture must be included.		
2	High Temperature	a) Temperature from 25 °C to 1200 °C with oven type surrounding heating to have uniform thermal gradient throughout the sample.		

	Stage (25 °C to 1200 °C)	Change from ambient to non-ambient should be computer controlled and user friendly.		
		b) Atmospheres: Air, inert gas, vacuum etc.		
		c) Provision to mount powder sample in the cavity holder and flat sample holder for solid samples.		
		d) Stage must have facility to rotate the sample to have improved statistics.		
		e) To avoid misalignment of sample height due to thermal expansion or contraction and deviation from X ray plane.		
		f) Quoted stage must have automatic Z(height) correction facility.		
		g) Quoted stage should have facility to be upgradable for mounting capillary holder for transmission geometry.		
		h) Fast heating and cooling		
		i) Accurate temperature measurement by a thermo sensor close to the sample		
		j) Easy handling and exchange of samples		
		k) Control unit, water Inlet/outlet connection.		
		l) Compact design		
		m) Software controlled operation must be available		
3.	Additional Mo Target to minimize fluorescence			
4	Additional Co Target to minimize fluorescence			
Important notes				
a.	The quoted XRD model should be upgradable to /compatible with	I. Residual Stress and Texture Measurement Facilities,		
		II. High Temperature Stage (25 °C to 1200 °C), and		
		III. Replacement of Cu target by Mo or Co target to minimize fluorescence		
b.	Depending on the fund available for the XRD, MNIT Jaipur will compare the total price of basic XRD unit plus any one or more optional items. Hence, the bidders should quote the best competitive prices for the basic XRD as well as optional items.			
c.	If technical committee asks at the time of evaluating technical specifications or before placing purchase order, the vendor should make a provision for demonstration on the model quoted or similar type of model either at their customer site(s) or service center.			
d.	Vendor should provide a declaration on letter head from the OEM for supplying all necessary accessories and replacement parts for next 10 years.			
5.	<u>Terms and conditions:</u>	a) Supply should be made within 12 weeks of the release of the purchase order.		
		b) Minimum 1-year warranty period is required. Vendors should provide continuous technical support and maintenance of the equipment during the warranty period.		
		c) The OEM or the vendor representing an OEM must have at least 5 years of experience in manufacturing/marketing such high-resolution optical equipment.		
		d) Vendors must have sufficient experience in supplying equipment to reputed organizations for research purposes. The experience of the end users will also be used as a criterion for the selection of bids that meet technical requirements. List of reputed end users inclusive of educational institutions in India to whom supplies are made with client contact details should be furnished. Relevant PO details are to be submitted as supporting documents.		
		e) All equipment must be compatible with Indian electrical standards and codes.		
		f) Cost breakup for all modules included in the scope of supply is mandatory. Optional items are to be quoted separately.		
		g) Non-OEM suppliers must submit a tender specific authorization from the OEM for supply to MNIT Jaipur.		
		h) Vendors must provide detailed documentation for the equipment.		

4. **GENERAL TERMS & CONDITIONS**

4.1. Due date: The tender has to be submitted on or before the due date. The offers received after the due date and time will not be considered.

4.2. Earnest Money Deposit (EMD): The Bidder should submit EMD as per “**Schedule**”. The Technical Bid without EMD would be considered as UNRESPONSIVE and will not be accepted. The EMD will be refunded without any interest to the unsuccessful Bidders after the award of contract. **The bidders are required to submit their bank details/cross cancelled cheque in this regard.** In case of successful bidder, it will be refunded after receipt of Performance Bank Guarantee. **NSIC / MSME registered bidders must submit copy of valid NSIC / MSME Registration Certificate for exemption of EMD.**

4.3. Opening of the tender: The bid will be opened by a committee duly constituted for this purpose. The bids (complete in all respect) received will be opened in presence of bidder’s representative if available. Only one representative will be allowed to participate in the tender opening. The technical bid will be opened first and it will be examined by a technical committee (as per specification and requirement). The financial offer/bid will be opened only for the offer/bid which technically meets all requirements as per the specification, and will be opened subsequently to determine the lowest quoted bid and purchase order will be awarded to firm (L₁) quoting lowest price.

4.4. Preference to Make in India Products:

The Institute is following and abide with the revised Public Procurement (Preference to Make in India), Order 2017 P- 45021/2/2017 – B. E. –II dated 04.06.20 issued by DPIIT, Ministry of Commerce and Industry, Govt. of India & subsequent amendments/instructions of Ministry. Accordingly, preference will be given to the make in India products while evaluating the bids. However, it is sole responsibility of the bidder(s) to specify the product quoted by them is of Make in India along with respective documentary evidence in the technical bid itself.

4.5. Purchase preference to Micro and Small Enterprises (MSEs):

Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference for product/services, the bidder must be the manufacturer/Service provider of the offered product/service. Relevant documentary evidence in this regard shall be submitted along with the bid in respect of the offered service. If L-1 is not an MSE and MSE Service Provider (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, then 100% order quantity will be awarded to such MSE bidder subject to acceptance of L1 bid price.

4.6. Acceptance/ Rejection of bids: The Institute reserves the right to reject any bid not fulfilling the eligibility criteria. Submission of incomplete bid/ incomplete bid format would lead to rejection of bids. All documents required to be submitted should be the part of the bid. If any document is not submitted, the bid will be treated as incomplete and this would lead to rejection. No communication in this regard will be entertained. Non-compliance of tender terms, non-submission of required documents, lack of clarity of the specifications, contradiction between bidder’s specification and supporting documents etc. may lead to rejection of the bid.

4.7. Pre Qualification Criteria:

- 4.7.1. Bidder should be the manufacturer / authorized dealer. Letter of Authorization from Original Equipment Manufacturer (OEM) specific to the tender should be enclosed in the format **(Annexure V)**. **Letter of authorization from OEM in any other format or not specific to this tender will not be accepted.**
- 4.7.2. An undertaking from the OEM is required stating that they would facilitate the bidder on a regular basis with technology/product updates and extend support for the warranty as well. **(Annexure II)**
- 4.7.3. In the tender, either the Indian agent on behalf of the Principal/OEM or Principal/OEM itself can bid but both cannot bid simultaneously for the same item/product in the same tender.
- 4.7.4. If an agent submits bid on behalf of the Principal/OEM, the same agent shall not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product.
- 4.7.5. The Bidder must have successfully executed / completed at least one single order of Rs. 96,00,000.00 or 2 orders each of Rs.60,00,000.00 or 3 orders each of Rs. 48,00,000.00 for similar products in last three years to any IIT/NIT/CFTIs/ any other Govt. organization. Copies of contracts / purchase orders and documentary evidence of successful execution / completion in support of Past Experience of Similar product along with names, address and contact details of clients shall be submitted in the format **(Annexure III)** with the bid for verification by the Buyer.
- 4.7.6. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be Rs.96,00,000.00. **Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be submitted with the bid.** In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for these criteria.
- 4.7.7. Class-I, Class-II local supplier and Non local supplier will be eligible to bid notified vide (DPIIT) Notification No. P- 45021/2/2017-PP (BE-II) dated 4th June, 2020. **(Submit duly filled Annexure VIII for the same)**

4.8. Performance Security: The supplier shall require to submit the performance security in the form of irrevocable bank guarantee issued by any commercial bank in the prescribed format **(Annexure VI)** for an amount which is stated at the "Schedule" of the tender document within 10 days from the date of receipt of the purchase order/LC and should be kept valid for a period of 60 days beyond the date of completion of warranty period.

4.9. Force Majeure: The Supplier shall not be liable for forfeiture of its performance security, liquidated damages or termination for default, if and to the extent that, it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.

- 1) For purposes of this Clause, "Force Majeure" means an event beyond the control of the Supplier and not involving the Supplier's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the Purchaser either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.
- 2) If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.

4.10. Risk Purchase Clause: In event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from the other source on the total risk of the supplier under risk purchase clause.

4.11. Packing Instructions: Each package will be marked on three sides with proper paint/indelible ink, the following:

- a) Item Nomenclature
- b) Order/Contract No.
- c) Country of Origin of Goods
- d) Supplier's Name and Address
- e) Consignee details
- f) Packing list reference number

4.12. Delivery and Documents:

Delivery of the goods should be made within a maximum period mentioned at the '**Schedule**' from the date of placement of purchase order. Within 24 hours of shipment, the supplier shall notify the purchaser and the insurance company by cable/telex/fax/e-mail the full details of the shipment including contract number, railway receipt number/ AAP etc. and date, description of goods, quantity, name of the consignee, invoice etc. The supplier shall submit the following documents to the purchaser:

- a) 4 Copies of the Supplier invoice showing contract number, goods' description, quantity
- b) unit price, total amount;
- c) Insurance Certificate if applicable;
- d) Manufacturer's/Supplier's warranty certificate;
- e) Inspection Certificate issued by the nominated inspection agency, if any
- f) Supplier's factory inspection report; and
- g) Certificate of Origin (if possible by the beneficiary);
- h) Two copies of the packing list identifying the contents of each package.
- i) The above documents should be received by the Purchaser before arrival of the Goods (except where the Goods have been delivered directly to the Consignee with all documents) and, if not received, the Supplier will be responsible for any consequent expenses.

4.13. Liquidated Damages (L.D):

If a supplier fails to execute the order in time as per the terms and conditions stipulated therein, it will be open to the purchaser to recover liquidated damages for delay in delivery and installation from the supplier at the rate 0.5% of the value of the order per week subject to a maximum of 10% of the total order value. The L.D charges can be increased in case of gross violation of the Purchase Order terms as decided by the Competent Authority of the Institute.

4.14. Prices: The price should be quoted in prescribed format (**Annexure IX**) only. The offer/bid should be exclusive of taxes and duties, which will be paid by the purchaser as applicable. However, the percentage of taxes & duties shall be clearly indicated. The price should be quoted without custom duty and excise duty, Since MNIT Jaipur is eligible for concessional rate of custom duty. Necessary certificate will be issued on demand.

4.15. Progress of Supply: Wherever applicable, supplier shall regularly intimate progress of supply, in writing, to the Purchaser as under:

- a) Quantity offered for inspection and date;
- b) Quantity accepted/rejected by inspecting agency and date;
- c) Quantity dispatched/delivered to consignees and date;
- d) Quantity where incidental services have been satisfactorily completed with date;

- e) Quantity where rectification/repair/replacement effected/completed on receipt of any communication from consignee/Purchaser with date;
- f) Date of completion of entire Contract including incidental services, if any; and
- g) Date of receipt of entire payments under the Contract (In case of stage-wise inspection, details required may also be specified).

4.16. Resolution of Disputes: The dispute resolution mechanism would be as follows:

- 4.16.1. In case of Dispute or difference arising between the Purchaser and a domestic supplier relating to any matter arising out of or connected with this agreement, such disputes or difference shall be settled in accordance with the Indian Arbitration & Conciliation Act, 1996, the rules thereunder and any statutory modifications or re-enactments thereof shall apply to the arbitration proceedings. The dispute shall be referred to the Director, MNIT Jaipur and if he is unable or unwilling to act, the sole arbitration of some other person appointed by him willing to act as such Arbitrator. The award of the arbitrator so appointed shall be final, conclusive and binding on all parties to this order.
- 4.16.2. In the case of a dispute between the purchaser and a Foreign Supplier, the dispute shall be settled by arbitration in accordance with provision of sub-clause (i) above. But if this is not acceptable to the supplier then the dispute shall be settled in accordance with provisions of UNCITRAL (United Nations Commission on International Trade Law) Arbitration Rules.
- 4.16.3. The venue of the arbitration shall be the place from where the order is issued.

4.17. Place of Jurisdiction: The place of jurisdiction would be Jaipur (Rajasthan).

4.18. Right to Use Defective Goods

If after delivery, acceptance and installation and within the warranty period, the operation or use of the goods proves to be unsatisfactory, the Purchaser shall have the right to continue to operate or use such goods until rectifications of defects, errors or omissions by repair or by partial or complete replacement is made without interfering with the Purchaser's operation.

4.19. Transfer and Subletting: The supplier shall not sublet, transfer, assign or otherwise part with the acceptance to the tender or any part thereof, either directly or indirectly, without the prior written permission of the Purchaser.

4.20. Supplier Integrity

The Supplier is responsible for and obliged to conduct all contracted activities in accordance with the Contract using state of the art methods and economic principles and exercising all means available to achieve the performance specified in the contract.

4.21. Installation & Demonstration

The supplier is required to do the installation and demonstration of the equipment within two weeks of the arrival of materials at the MNIT Jaipur, site of installation; otherwise the penalty clause will be the same as per the supply of materials.

In case of any damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. MNIT Jaipur will not be liable to any type of losses in any form.

4.22. Insurance (if applicable): For delivery of goods at the purchaser's premises (MNIT Jaipur), the insurance shall be obtained by the supplier from "warehouse to warehouse" (final destinations) on "All Risks" basis including War Risks and Strikes. The insurance shall be valid for a period of not less than 3 months after installation and commissioning.

4.23. Warranty:

- 4.23.1. Warranty period shall be (as stated at “Schedule “of this tender) from date of installation of Goods and acceptance at MNIT Jaipur. The Supplier shall, in addition, comply with the performance and/or consumption guarantees specified under the contract. If for reasons attributable to the Supplier, these guarantees are not attained in whole or in part, the Supplier shall at its discretion make such changes, modifications, and/or additions to the Goods or any part thereof as may be necessary in order to attain the contractual guarantees specified in the Contract at its own cost and expense and to carry out further performance tests. The warranty should be comprehensive on site.
- 4.23.2. The Purchaser shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall arrange to repair or replace the defective goods or parts within 3 days free of cost in MNIT Jaipur. The Supplier shall take over the replaced parts/goods at the time of their replacement. No claim whatsoever shall lie on the Purchaser for the replaced parts/goods thereafter. The period for correction of defects in the warranty period is 03 days. If the supplier, having been notified, fails to remedy the defects within 03 days, the purchaser may proceed to take such remedial action as may be necessary, at the supplier’s risk and expenses and without prejudice to any other rights, which the purchaser may have against the supplier under the contract.
- 4.23.3. The warranty period should be clearly mentioned. The maintenance charges (AMC) under different schemes after the expiry of the warranty should also be mentioned. The comprehensive warranty will commence from the date of the satisfactory installation/commissioning of the equipment against the defect of any manufacturing, workmanship and poor quality of the components.

4.24. Governing Language

The contract shall be written in English language. English language version of the Contract shall govern its interpretation. All correspondence and other documents pertaining to the Contract, which are exchanged by the parties, shall be written in the same language.

4.25. Applicable Law

The Contract shall be interpreted in accordance with the laws of the Union of India and all disputes shall be subject to place of jurisdiction.

4.26. Notices

- 4.26.1. Any notice given by one party to the other pursuant to this contract/order shall be sent to the other party in writing or by cable, telex, FAX or e-mail and confirmed in writing to the other party’s address.
- 4.26.2. A notice shall be effective when delivered or on the notice’s effective date, whichever is later.

4.27. Taxes and Duties

Suppliers shall be entirely responsible for all taxes, duties, license fees, octroy, road permits, etc., incurred until delivery of the contracted Goods to the Purchaser. However, GST in respect of the transaction between the Purchaser and the Supplier shall be payable extra, if so stipulated in the order.

4.28. Payment:

- 1) For Indigenous supplies, 100% payment shall be made by the Purchaser against delivery, inspection, successful installation, commissioning and acceptance of the equipment at MNIT Jaipur in good condition and to the entire satisfaction of the Purchaser and on production of unconditional performance bank guarantee as specified in bid document.

- 2) For imported items Payment will be made through irrevocable Letter of Credit (LC). Letter of Credit (LC) will be established in favor of foreign Supplier after the submission of performance security. The letter of credit (LC) will be established on the exchange rates as applicable on the date of establishment. For Imports, LC will be opened for 100% CIF/CIP, Delhi value. 80% of the LC amount shall be released on presentation of complete and clear shipping documents and 20% of the LC amount shall be released after the installation and demonstration of the equipment at the MNIT Jaipur site of installation in faultless working condition for period of 60 days from the date of the satisfactory installation and subject to the production of unconditional performance bank guarantee as specified in as specified in bid document.
- 3) GST Deduction at source as per Order/ notification of the Govt.
- 4) GST No of MNIT Jaipur is **08AAAJM0351L1Z6**
- 5) HSN/SAC No of the items must be clearly mentioned in the quotation along with GST No.

4.29. Manuals and Drawings:

- 4.29.1. Before the goods and equipment are taken over by the Purchaser, the Supplier shall supply operation and maintenance manuals. These shall be in such details as will enable the Purchaser to operate, maintain, adjust and repair all parts of the works as stated in the specifications.
- 4.29.2. The Manuals shall be in the ruling language (English) in such form and numbers as stated in the contract.
- 4.29.3. Unless and otherwise agreed, the supply of goods shall not be considered to be completed for the purposes of taking over until such manuals and drawing have been supplied to the Purchaser.

4.30. After Sale Service certificate: The bidder must mention the details of 'After Sale Services' in the format (**Annexure- IV**). **Without After sale service certificate in the prescribed format, bid will not be considered technically responsive.**

4.31. Application Specialist: The Tenderer should mention in the Techno-Commercial bid the availability and names of Application Specialist and Service Engineers in the nearest regional office. (Ref. to Annexure- III)

4.32. Site Preparation: The supplier shall inform to the Institute about the site preparation, if any, needed for the installation of equipment, immediately after the receipt of the purchase order. The supplier must provide complete details regarding space and all the other infrastructural requirements needed for the equipment, which the Institute should arrange before the arrival of the equipment to ensure its timely installation and smooth operation thereafter. The supplier may visit the Institute and see the site where the equipment is to be installed and may offer his advice and render assistance to the Institute in the preparation of the site and other pre- installation requirements.

4.33. Spare Parts: The Supplier may be required to provide any or all of the following materials, notifications, and Information pertaining to spare parts manufactured or distributed by the Supplier. Such spare parts as the Purchaser may elect to purchase from the Supplier, providing that this election shall not relieve the Supplier of any warranty obligations under the Contract; and In the event of termination of production of the spare parts; Advance notification to the Purchaser of the pending termination, in sufficient time to permit the Purchaser to procure needed requirements; and following such termination, furnishing at no cost to the Purchaser, the blueprints, drawings and specifications of the spare parts, if requested. Supplier shall carry sufficient inventories to assure ex-stock supply of

consumable spares. Other spare parts and components shall be supplied as promptly as possible but in any case within six months of placement of order.

- 4.34. Defective Equipment:** If any of the equipment supplied by the Supplier is found to be substandard, refurbished, un-merchantable or not in accordance with the description/specification or otherwise faulty, the committee will have the right to reject the equipment or its part. The prices of such equipment shall be refunded by the Supplier with 18% interest if such payments for such equipment have already been made. All damaged or unapproved goods shall be returned at suppliers cost and risk and the incidental expenses incurred thereon shall be recovered from the supplier. Defective part in equipment, if found before installation and/or during warranty period, shall be replaced within 7 days on receipt of the intimation from this office at the cost and risk of supplier including all other charges. In case supplier fails to replace above item as per above terms & conditions, MNIT Jaipur may consider "Banning" the supplier.
- 4.35. Termination for Default:** The Purchaser may, without prejudice to any other remedy for breach of contract, by written notice of default sent to the Supplier, terminate the Contract in whole or part:
- i. If the Supplier fails to deliver any or all of the Goods within the period(s) specified in the order, or within any extension thereof granted by the Purchaser; or
 - ii. If the Supplier fails to perform any other obligation(s) under the Contract.
 - iii. If the Supplier, in the judgment of the Purchaser, has engaged in corrupt or fraudulent practices in competing for or in executing the Contract.
 - iv. For the purpose of this Clause:
 1. "Corrupt practice" means the offering, giving, receiving or soliciting of anything of value to influence the action of a public official in the procurement process or in contract execution.
 2. "Fraudulent practice" means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of the Borrower, and includes collusive practice among bidder (prior to or after bid submission) designed to establish bid prices at artificial non-competitive levels and to deprive the Borrower of the benefits of free and open competition;”
 - v. In the event the Purchaser terminates the Contract in whole or in part, the Purchaser may procure, upon such terms and in such manner, as it deems appropriate, Goods or Services similar to those undelivered, and the Supplier shall be liable to the Purchaser for any excess costs for such similar Goods or Services. However, the Supplier shall continue the performance of the Contract to the extent not terminated.
- 4.36. Training of Personnel (if applicable):** The supplier shall be required to undertake to provide the technical training to the personnel involved in the use of the equipment at the Institute premises, immediately after completing the installation of the equipment if applicable.
- 4.37. Downtime:** During the warranty period, not more than 1% downtime will be permissible. For every day exceeding permissible downtime, penalty of 1/365 of the 1% contract value will be imposed. Downtime will be counted from the date and time of the filing of complaint within the business hours. Supplier should clearly mention about their service set up in India (preferably in Northern part of India) for prompt service support. The manufacturer and/or their Indian representative must have at least three qualified and factory trained service engineer in India to be able to attend to service at MNIT Jaipur within 48 hours on submitting a complaint. Training

certificates from the manufacturer have to be provided with the tender. During the warranty period, only factory trained and certified engineers are acceptable to attend the service. In case the Equipment / System remains non-operational for more than 5 days then warranty period will be extended for the equivalent period for which Equipment /System remained non- operational. Warranty extension in such case shall be done without prejudice to any other Term &condition of the contract. MNIT Jaipur would like to enter in service agreement through which MNIT Jaipur will receive replacement of defective spares/part (if any, that are not covered under warranty) immediately so as to minimize the down time. Order, if any, required to be placed for such spares/parts will be done by MNIT Jaipur in due course of time.

4.38. Compliance certificate: This certificate must be provided indicating conformity to the technical specifications. (**Annexure-I**)

4.39. Genuine Pricing: Vendor is to ensure that quoted price is not more than the price offered to any other customer in India to whom this particular item has been sold.

4.40. Comparison of Bids:

Only the short-listed bids from the technical evaluation shall be considered for commercial comparison.

For comparison of prices, the prices quoted by the bidder for all the items taken together, as per Price Bid shall be considered.

The bids shall be evaluated on the basis of the total prices including all taxes and duties.

For the purpose of comparison of prices on equal basis, the prices quoted in foreign currency (if any) by bidders shall be converted to equivalent Indian Rupees. The exchange rate as on the date of opening of financial bids shall be considered for the purpose of comparison.

4.41. Award of Contract

MNIT Jaipur shall award the contract to the eligible bidder whose technical bid has been accepted and determined as the lowest evaluated commercial bid based on the criteria mentioned above. However, MNIT Jaipur reserves the right and has sole discretion to reject the lowest evaluated bid.

In case more than one bidder happens to quote the same lowest price, MNIT Jaipur reserves the right to decide the criteria and further process for awarding the contract. Decision of MNIT Jaipur shall be final for awarding the contract.

4.42. Restriction on procurement from bidder of country sharing land border with India:

As per Ministry of Finance, Deptt. of Expenditure, Public Procurement Division Order (Public Procurement No.1) issued from file No.6/18/2019-PPD dated 23rd July, 2020 regarding restrictions under Rule 144 (xi) of the General Financial Rules (GFRs) 2017, a certificate (**Annexure VII**) shall be submitted by bidders in the tender documents regarding their compliance with the said order. If the certificate submitted by a bidder whose bid is accepted is found to be false, this would be a ground for immediate termination and further legal action in accordance with law.

4.43. Exemption to Startups:

If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria". In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

4.44. Exemption to MSME:

If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria". In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

4.45. Bidder must enclose the duly signed and stamped tender document.

Deputy Registrar
Stores and Purchase MNIT Jaipur

COMPLIANCE SHEET

Name of the Item : Powder X-ray Diffraction with Residual Stress Measurement Facility/ Quantity: 1			
No.			
SI No.	Particulars	Comply (Yes/No)	Supporting Page No.
1	Floor mounted fully automated X-Ray Diffraction (XRD) system along with all necessary hardware and software for analysis of powder and solid samples of metals/alloys/ceramic/polymer/composites. In future the system must have capability to be upgradable for SAXS, Microdiffraction, Texture and In-Situ Non-Ambient/high temperature Applications.		
3.	X-Ray Generator		
	f) Output power: 3 kW or more		
	g) Voltage: up to 60kV or higher with increment of 1 kV		
	h) Anode Current: up to 50mA or higher with increment of 1 mA		
	i) Control: Fully automated through software		
	j) Stability: ≤ 0.01 % per 10 % variation of main voltage or better at 50 Hz		
4.	Goniometer		
	h) Type: Higher resolution Theta/Theta		
	i) Scan: θ/θ coupled and θ/θ decoupled. Goniometer should be with independent θ and 2θ drives for accurate positioning and with design to have zero backlash.		
	j) Two theta range: from -90° to 162° or better range		
	k) Minimum step size: 0.0001° or better		
	l) Angular speed: 10-15 degree/sec or better		
	m) Angular resolution: 0.028 deg. or better		
	n) Goniometer radius: 220 mm or more. Suitable hardware and software to be included for goniometer radius change facility.		
3.	X-Ray Tube		
	i) Target: Copper (Cu)		
	j) Focus: Long fine focus (LFF) X-ray tube with 1 line and 1 point focus.		
	k) Insulation: Ceramic Insulation.		
	l) Maximum output power: ≥ 1.8 kW		
	m) Suitable Beta filter to be included.		
	n) X-ray tube must have auto component recognition facility to indicate anode target type and its exchange from one source type to another type.		
	o) Changeover from line to point focus and Vice-Versa should be available.		
	p) Alignment of X-Ray tube Height and Focus must be through computer control.		
4.	Sample stage and Sample holder		
	c) Standard spinning/rotating sample stage(s) suitable for reflection and transmission mode diffraction with facility to keep powder sample always in horizontal position.		
	d) For solid and thin film sample, a suitable sample stage to be offered with computer controlled motorized Z stage with up to 25 mm or more movement and capability to mount sample up to 2kg weight. The sample stage must be able to handle sample of minimum 100 mm diameter and 50 mm thickness		
5.	Optics		
	All optical components and its settings must be automatically recognized, changed and aligned by XRD control software for error free analysis with latest design.		
	h) Primary Beam Optics: Computer controlled motorized divergence slits, two numbers of suitable sollers slit/ s, motorized mask and beta filters to be offered. Multilayer mirror for Cu X-ray source to be offered for thin film and transmission. Changeover from Bragg Brentano geometry to parallel beam geometry must be computer controlled.		
	i) Secondary Beam: Computer controlled motorized Anti-scatter slits, suitable soller slit/s, long soller slit/parallel plate collimator to be offered for parallel beam geometry.		
	j) Filters: Suitable $K\beta$ -filters for Cu radiation		
	k) Monochromator to reduce the fluorescence from the sample.		
	l) The Line optics must include X-ray Mirror to carry out, thin film GIXRD.		

		m) For carrying out residual stress analysis, point focus of the x-ray beam must be useful.		
		n) Necessary beam collimator to be offered with diameter of 2mm, 0.01 mm.		
6.	X-ray Detector	The maintenance free solid state (semiconductor) detector should have following specifications;		
		j) High degree of linearity (i.e., 95% in the range of 0 to 10 ⁸ CPS)		
		k) Very less background noise		
		l) 1D detector should be capable of working in 0D and 1D mode.		
		m) Number of active channels / pixels 190 or higher.		
		n) Spatial resolution should be 75 micron or less.		
		o) Detector should work in continuous, fast scan, slow scan and static modes.		
		p) It should also work in fluorescence reduction mode.		
		q) The detector should be able to work with most of the X-ray source like Cu, Mo, Co etc.		
		r) The detector should not use any liquid or gas in its operation.		
7.	Standards	d) Fluorescent disc/screen for positioning the X-ray beam		
		e) Silicon and/or other standard sample for checking the accuracy of the peak position within + 0.01 linearity range over the entire goniometer angular range.		
		f) Reference standard for Residual Stress Application		
8.	Software	f) Data collection - Licensed version of the software to control the complete XRD instrument, auto component recognition of all the components, system diagnosis and programming for data collection to be offered.		
		g) Data analysis - Licensed version of data evaluation software for background correction, profile fitting, search match, qualitative analysis, structure refinement, lattice parameter refinement, crystallite size determination, Rietveld based standardless quantification refinement package with Structure analysis, profile fitting, lattice parameter refinement, stripping of K α 2 peak, and crystallite size determination software package should be offered. Qualitative and quantitative analysis software, search match, dedicated SAXS application software for analysis of bulk and nanoparticle has to be offered.		
		h) Latest version of ICDD PDF4+ AXIOM Database with minimum 3 years validity must be quoted.		
		i) Software for data collection and data analysis of thin film GIXRD analysis and residual stress analysis to be offered.		
		j) Data analysis software should be with minimum 3 user licenses for use by multiple scientists, faculties and researchers.		
9	XRD-analysis Computer	Suitable high end computer with all the software (as mentioned in point 8), 23inch monitor and a device to produce colored XRD patterns.		
10.	Safety:	d) Provision for detection of abnormal cooling water flow, pressure and temperature, abnormal generator overload and shutter malfunction detection.		
		e) XRD system should be compliant as per International Safety Standards and Regulation norms like Eurotom, pertaining mechanical, electrical and X-ray Radiation safety and other hazards.		
		f) The supplier must have necessary Type approval from AERB for supply of XRD and necessary approval document to be submitted along with offer.		
11	Cooling system:	c) A suitable external water chiller compatible with the main instrument whose Make and detailed specifications to be provided with the installation requirements. They should preferably be domestically procured.		
		d) External filter/traps should be provided for minimizing the dust settlements from the chiller to the target area.		

12	Power backup system for XRD	c) Minimum 15 kVA or more IGBT based online power backup with built-in isolation transformer.		
		d) Power back must be able to handle XRD system, XRD-analysis Computer, and external water chiller pump for more than 30 min uninterrupted XRD data collection during main power failure.		
13	Other mandatory requirements	f) List of recent installations (in last 3 years) of the offered XRD model in India should be provided.		
		g) After sales service will be a component for technical evaluation of the equipment. For this, list of service centers in north region with contact details of engineers must be submitted.		
		h) XRD system should be supported by manufacturer for the supply of spares for at least 8-10 years after Installation.		
		i) All pre-installation requirements such as room size, mechanical vibration and required power rating for the XRD system and water chiller are to be provided. The supplier must arrange site inspection for installation planning. The vendor should provide all required accessories and tackle for complete installation and commissioning of instrument.		
		j) All the operational manual, application manual and service manual along with schematic in English are to be provided. Test Reports for all the modes of operation to be provided.		
14	Application and Operation Training:	c) Supplier must provide complete operation, DO'S AND DON'TS, and application training for all aforesaid applications for minimum 5 days.		
		d) All costs towards training including lodging, boarding and travel have to be borne by the XRD supplier.		
15	Warranty:	d) XRD system including all accessories such as X-Ray Generator, X-ray tube, electronics, motors and detector must be provided with complete & comprehensive warranty of 3 years.		
		e) During warranty period, if any part is required for replacement, then same to be arranged on delivered duty paid (DDP) basis without any cost to buyer.		
		f) X-ray tube should be supplied before expiry of 3 rd year. This is to make sure that the XRD system will work without interruption for 5 years.		
Optional Item				
1	Accessories for residual stress and texture measurement	e) 3 Axis Eulerian cradle with Chi and Phi movement with automatic computer control to be included.		
		f) Advance polycapillary lens to have point focus beam on sample for stress and texture measurement.		
		g) Sample holder adaptor for cradle must be provided.		
		h) Necessary software for texture must be included.		
2	High Temperature Stage (25 °C to 1200 °C)	n) Temperature from 25 °C to 1200 °C with oven type surrounding heating to have uniform thermal gradient throughout the sample. Change from ambient to non-ambient should be computer controlled and user friendly.		
		o) Atmospheres: Air, inert gas, vacuum etc.		
		p) Provision to mount powder sample in the cavity holder and flat sample holder for solid samples.		
		q) Stage must have facility to rotate the sample to have improved statistics.		
		r) To avoid misalignment of sample height due to thermal expansion or contraction and deviation from X ray plane.		
		s) Quoted stage must have automatic Z(height) correction facility.		
		t) Quoted stage should have facility to be upgradable for mounting capillary holder for transmission geometry.		
		u) Fast heating and cooling		
		v) Accurate temperature measurement by a thermo sensor close to the sample		
		w) Easy handling and exchange of samples		
		x) Control unit, water Inlet/outlet connection.		
	y) Compact design			

		z) Software controlled operation must be available		
3.	Additional Mo Target to minimize fluorescence			
4	Additional Co Target to minimize fluorescence			
Important notes				
a.	The quoted XRD model should be upgradable to /compatible with	IV. Residual Stress and Texture Measurement Facilities,		
		V. High Temperature Stage (25 °C to 1200 °C), and		
		VI. Replacement of Cu target by Mo or Co target to minimize fluorescence		
b.	Depending on the fund available for the XRD, MNIT Jaipur will compare the total price of basic XRD unit plus any one or more optional items. Hence, the bidders should quote the best competitive prices for the basic XRD as well as optional items.			
c.	If technical committee asks at the time of evaluating technical specifications or before placing purchase order, the vendor should make a provision for demonstration on the model quoted or similar type of model either at their customer site(s) or service center.			
d.	Vendor should provide a declaration on letter head from the OEM for supplying all necessary accessories and replacement parts for next 10 years.			
5.	Terms and conditions:	i) Supply should be made within 12 weeks of the release of the purchase order.		
		j) Minimum 1-year warranty period is required. Vendors should provide continuous technical support and maintenance of the equipment during the warranty period.		
		k) The OEM or the vendor representing an OEM must have at least 5 years of experience in manufacturing/marketing such high-resolution optical equipment.		
		l) Vendors must have sufficient experience in supplying equipment to reputed organizations for research purposes. The experience of the end users will also be used as a criterion for the selection of bids that meet technical requirements. List of reputed end users inclusive of educational institutions in India to whom supplies are made with client contact details should be furnished. Relevant PO details are to be submitted as supporting documents.		
		m) All equipment must be compatible with Indian electrical standards and codes.		
		n) Cost breakup for all modules included in the scope of supply is mandatory. Optional items are to be quoted separately.		
		o) Non-OEM suppliers must submit a tender specific authorization from the OEM for supply to MNIT Jaipur.		
		p) Vendors must provide detailed documentation for the equipment.		
6	Proof of Earnest money Deposit			
7	Copy of GST Certificate and PAN			
8	Leaflets catalogue (Brochure)			
9	Non Blacklisting of Supplier and Price reasonability declaration			
10	Required past Experience			
11	Minimum average annual turnover of the bidder			
12	After Sale Service Certificate			
13	Manufacturer's Authorization Letter			
14	Certificate - Bidder Not from/ from Country sharing Land border with India & Registration of Bidder with Competent Authority			
15	Declaration of Local Content			

Signature of Bidder

Name: _____

Designation: _____

Organization Name:

Contact No.: _____

<< Organization Letter Head >>

DECLARATION SHEET

We, _____ hereby certify that all the information and data furnished by our organization with regard to these tender specifications are true and complete to the best of our knowledge. I have gone through the specifications, conditions and stipulations in details and agree to comply with the requirements and intent of specification.

This is certified that our organization has been authorized (Copy attached) by the OEM to participate in Tender. We further certify that our organization meets all the conditions of eligibility criteria laid down in this tender document. Moreover, OEM has agreed to support on regular basis with technology / product updates and extend support for the warranty.

We, further specifically certify that our organization has not been Black Listed/De Listed or put to any Holiday by any Institutional Agency/ Govt. Department/ Public Sector Undertaking in the last three years.

The prices quoted in the financial bids are subsidized due to academic discount given to MNIT Jaipur and the rates quoted are not more than those quoted to any other Institution in India or abroad during the last one year.

Signature of Bidder

Name: _____

Designation: _____

Organization Name:

Contact No.: _____

AFTERSALE SERVICE CERTIFICATE

From:

To

The Registrar,
 Malaviya National Institute of Technology (MNIT),
 Jaipur

Whereas, we M/s (Bidder Name) are established & reputable manufacturers (Make of items) of [items name] having service offices at Delhi, Jaipur and in the state of Rajasthan. Details are as under:

Sr.No. Address of Service Centre Phone No. Number of Engineers

- 1.
- 2.
- 3.

 We do hereby confirm that:

Services including repair/replacement of defective parts will be done by us. Replacement of defective Systems/parts will be done by equivalent or better systems/parts of the same make. We will attend all the complaints/service calls within 24 working hours and not beyond 3 working days. Down time will not exceed beyond 3 working days. In case, down time exceed 3 working days then we will extend the warranty period of that item(s) double of the down time.

Signature of Bidder**Name:** _____**Designation:** _____**Organization Name:** _____**Contact No.:** _____

MANUFACTURERS' AUTHORIZATION FORM

[The bidder shall require the Manufacturer to fill in this Form in accordance with the Instructions indicated. This letter of authorization should be on the letter head of the Manufacturer and should be signed by a person with the proper authority to sign documents that are binding on the Manufacturer]

Date: [insert date (as day, month and year) of Bid Submission] Tender No.: [insert number from Invitation for Bids]

To: [insert complete name and address of Purchaser] WHEREAS

We [insert complete name of Manufacturer], who are official manufacturers of [insert type of goods manufactured], having factories at [insert full address of Manufacturer's factories], do hereby authorize [insert complete name of Bidder] to submit a bid the purpose of which is to provide the following Goods, manufactured by us [insert name and or brief description of the Goods], and to subsequently negotiate and sign the Contract.

We hereby extend our full guarantee and warranty in accordance with Clause 4.23 of the Terms and Conditions, with respect to the Goods offered by the above firm. Further We also ensure to support on regular basis with technology / product updates.

Signed: [insert signature(s) of authorized representative(s) of the Manufacturer]

Name: [insert complete name(s) of authorized representative(s) of the Manufacturer]

Title: [insert title]

Duly authorized to sign this Authorization on behalf of: [insert complete name of Bidder]

Dated on day of , [insert date of signing]

PERFORMANCE BANK GUARANTEE

(To be executed on Stamp Paper of Rs. 100/- or such higher value as per the Stamp Act of the State in which the Guarantee is issued. Stamp Paper should be in the name of the Bank Issuing the Guarantee.)

BANK GUARANTEE NO. :

DATED :

Dear Sirs,

1. THIS DEED OF GUARANTEE made on this day of..... between **MALAVIYA NATIONAL INSTITUTE OF TECHNOLOGY, JAIPUR** (hereinafter called the "MNIT" which expression shall unless excluded by or repugnant to the context includes its successors and assignees) of the one part and the (hereinafter called the "**Bank**" which expression shall unless excluded by or repugnant to the context include its successors and assignees) of the other part.
2. AND WHEREAS as per clause..... Of the purchase order in question the supplier shall furnish a Performance Bank Guarantee of 03% of P.O. Value i.e. Rs. (in words) only) valid for the period of two months beyond warranty period as and by way of security for satisfactory working of the AND WHEREAS at the request of the supplier, the Bank executes these presents.
- 3.0 THIS DEED WITNESSETH AND IT IS HEREBY AGREED AND DECLARED BY AND BETWEEN PARTIES HERETO AS FOLLOWS:
 - 3.1 The Bank hereby guarantees to the MNIT, Jaipur that the equipment / service contracted are capable of performing the work as demanded by the MNIT, Jaipur. In the event of equipment / service failing to perform to the satisfaction of the MNIT, Jaipur, which shall be final and conclusive of the factum of non-performance, the Bank shall indemnify and keep the indemnified to the extent of of P.O. Value i.e. Rs. (Rupees) valid for the period of two months beyond the warranty period against any loss or damage that may be caused to or suffered by the MNIT, Jaipur consequent to non-performance of the contracted equipment / services to be supplied by the supplier.
 - 3.2 In consideration of the aforesaid premise and at the request of the supplier, we the Bank hereby irrevocably and unconditionally guarantee that the supplier shall perform in an orderly manner their contractual obligations in accordance with the terms and conditions set forth in the Purchase order dated and in the event of the supplier's failure to do so, the Bank unconditionally pay to the MNIT, Jaipur on demand, any amount up to the value mentioned in Clause 3.1 above without any reference to the supplier and without questioning the claim.
 - 3.3 The guarantee herein shall remain in full force for a period of two months beyond the warranty period from the date of certification by the MNIT, Jaipur of successful installation and commissioning of the equipment/ service contracted. Date of start of warranty period will be notified by MNIT, Jaipur to the Bank.

- 3.4 The decision of the MNIT, Jaipur regarding the liability of the Bank under the guarantee and the amount payable there under shall be final and conclusive and binding on us without question. The Bank shall pay forthwith the amount demanded by the MNIT, Jaipur notwithstanding any dispute, if any, between the MNIT, Jaipur and the supplier.
- 3.5 The Bank further agrees that the guarantee herein shall remain in full force during the pendency of aforesaid period mentioned in Clause 3.3 above and also any extension of the guarantee which has been provided by the Bank for this purpose beyond the aforesaid period provided, further, that if any claim accrues or against the Bank by virtue of this guarantee, should be lodged with us within a period of 60 days from the date of expiry of the guarantee period.
- 3.6 This Guarantee shall not be affected by any change in constitution of the supplier, MNIT, Jaipur or us nor shall it be affected by any change in constitution or by any amalgamation or absorption or reconstruction thereof otherwise, but will ensure for and be available to and endorsable by the absorbing amalgamated company or concern.
- 3.7 The MNIT, Jaipur has the fullest liberty without affecting the guarantee to postpone at any time or from time any of the powers exercisable by it against the supplier, either to enforce or forbear the clause governing guarantee in the terms and conditions of the said contract and Bank shall not be released from its liabilities under the guarantee by any matter referred to or by reason of time being given to the supplier or any other forbearance, act or omission on the part of the MNIT, Jaipur or any material or things whatsoever which under the law relating to sureties shall but for the provisions hereof have the effect of so releasing the Bank from its liabilities.
- 3.8 We further agree that the MNIT, Jaipur shall have the fullest liberty without affecting in any way our obligations hereunder with or without our consent or knowledge to vary any of the terms and conditions of the said contract or to extend the time of delivery from time to time.
- 3.9 The Bank undertakes not to revoke this guarantee during its currency except with the previous consent in writing of the MNIT, Jaipur.
- 3.10 We further agree that in order to give full effect to the guarantee herein contained MNIT, Jaipur shall be entitled to act as if we were its principal debtors in respect of its claim against the Supplier hereby guaranteed by us as aforesaid and we hereby expressly waive all our rights of suretyship and other rights if any which are in any way inconsistent with the above provision of this Guarantee.

Notwithstanding anything herein before, liability of the Bank under this guarantee is restricted to Rs. (Rupees only) and it will remain in force up to the period specified in Clause 3.3 unless a suit to enforce any claim under the Guarantee is filed against the Bank before the period specified in Clause 3.4. All your rights under this Guarantee shall be forfeited and we shall be relieved and discharged from all liabilities thereunder.

COUNTERSIGNED

Signature	:	Signature	:
Name	:	Name	:
Designation	:	Designation	:
Organization	:	Organization	:

<On Organization Letter Head>

ANNEXURE-VII

No. _____

Dated: _____

CERTIFICATE

I have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and hereby certify that the organization is not from such a country.

OR (whichever is applicable)

I have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and hereby certify that the organization is from _____ (*Name of Country*) and has been registered with the Competent Authority. I also certify that the organization fulfills all the requirements in this regard and is eligible to be considered.

(Copy/ evidence of valid registration by the Competent Authority is to be attached)

Signature of Bidder/ Agent

Name: _____

Designation: _____

Organization Name: _____

Contact No. : _____

DECLARATION OF LOCAL CONTENT
[For Local Content of Products, Services or Works]

(To be given on Company Letter Head – For tender value below Rs.10 Crores)

(To be given by Statutory Auditor/Cost Auditor/Cost Accountant/CA for tender value above Rs.10 Crores)

To,

The Registrar
MNIT Jaipur

Subject: Declaration of Local Content

Tender reference No. _____

1. Country of Origin of Goods being offered: _____
2. With reference to Order No. P- 45021/2/2017-PP(BE-II) dated 16-09-2020 of DPIIT, Ministry of Commerce and Industry, Govt. of India, we fall under the following category of supplier (please tick the correct category) for the items for which this tender has been floated and being bided.
 - Class I local supplier – has local content equal to more than 50%. Local contents added at _____ (name of location).
 - Class II local supplier – has local content more than 20% but less than 50%. Local content sadded at _____ (name of location).
 - Non-local supplier – has local content less than or equal to 20%. Local contents added at _____ (name of location).
3. We are solely responsible for the above mentioned declaration in respect of category of supplier. False declarations will be in breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which we may be debarred for up to 2 years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.

Signature of Bidder/ Agent

Name: _____

Designation: _____

Organization Name: _____

Contact No. : _____

PRICE BID FORMAT

PRICE SCHEDULE FROM (IF GOODS BEING OFFERED FROM INDIA)

Name of Bidder

Tender No. F5(2736)ST/MNIT/Mett./2022

1	2	3	4	5	6	7	8	9	10	11	12
Sl. No.	Item Description with HSN Code	Country of Origin	Unit	Quantity	Unit Price Ex-Works, Ex-Warehouse, Ex-show room off the shelf price (inclusive of all taxes already paid) 5x6)	Total Price Ex-Works, Ex-warehouse, Ex-show room off the shelf price (inclusion of all taxes already paid) 5x6)	GST & other taxes payable, if contract is awarded	Packing & forwarding up to station of dispatch, if any	Charges for inland transportation, insurance up to lab./ instt. By air/road/rail (retain one only)	Total Price	Installation, commissioning and training charges, if any
1.											

Total bid price in Indian Currency*: (in figures) :
 (in words):

Signature of Bidder
 Name
 Business Address

PRICE SCHEDULE FROM (IF GOODS BEING OFFERED FROM ABROAD)

Name of Bidder

Tender No. F5(2736)ST/MNIT/Mett./2022

1	2	3	4	5	6	7	8	9	10	11
Sl. No.	Item Description	Country of Origin	Unit	Quantity	Unit Price indicating currency FOB (named port of shipment or FCA (named place of delivery) (retain only one	Total Price (5x6) FOB (named port of shipment or FCA (named place of delivery) (retain only one	Charges for insurance & transportation to port/place of destination	Total price CIF/CIP Delhi (retain only one) (7x8)	Approx. shipment weight and volume	Indian customs Tariff No and HSN No. (ICT & HSN No.)
1.										

Total bid price for CIP/CIF Delhi in Foreign Currency : (in figures) :
 (in words):

Format (A or B) is to be used as applicable.

Signature of Bidder
 Name